



2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com



ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030562-11	367274	≤ -10 °C	Acetone w/ 0.1% Formic Acid	20-Jan-2021	Oxyfluorfen Solution, 100 mg/l, 1 ml

<u>Compound</u>	<u>CAS No.</u>	<u>Purity (%)</u>	<u>Neat Material Lot No.</u>	<u>Concentration</u>
oxyfluorfen	42874-03-3	98.8	562.3.3P	100.8 ± 1.52 mg/l

Certified By: _____

Thomas Churchill

Manufacture Date 21-Jan-2019

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.
This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.
The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.

Certificate of Analysis

Page 2 of 1

Catalog No.: 030562-11

Lot No.: 367274

Expiration Date: 20-Jan-2021

Compound

CAS No.

Purity (%)

Neat Material Lot No.

Concentration



Certified By: _____

Thomas Churchill

Manufacture Date 21-Jan-2019

Follow all storage requirements, keep tightly closed when not in use and use good laboratory practices when handling.

This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.