



7290B Investment Drive North Charleston, South Carolina 29417  
Phone: 866.272.0932 Fax: 866.509.5146 www.o2si.com



ISO 17025 Accredited Chemical Testing Lab  
Cert. No. 3031.01

Date Received: \_\_\_\_\_

# Certificate of Analysis

Rev 0

Page 1 of 1

<b>Catalog No.:</b> 120687-02	<b>Lot No.:</b> 419030	<b>Storage:</b> ≤ -10 °C -5PAK	<b>Solvent:</b> P/T Methanol	<b>Exp. Date:</b> 20-Aug-2021	<b>Description:</b> Custom PCE/TCE Standard, 2000 mg/L, 5 x 1 ml
-------------------------------	------------------------	-----------------------------------	------------------------------	-------------------------------	--

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
tetrachloroethylene	127-18-4	100	183.1.2P	2005 ± 20.83 mg/L
trichloroethylene	79-01-6	98.1	188.29.1P	2001 ± 28.85 mg/L

Certified By: \_\_\_\_\_

Erica Lawson

Manufacture Date 31-Jul-2020

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.

*This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.*

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.

# Certificate of Analysis

Catalog No.: 120687-02-5PAK

Lot No.: 419030

Expiration Date: 20-Aug-2021

Compound

CAS No.

Purity (%)

Neat Material Lot No.

Concentration



Certified By: \_\_\_\_\_

Erica Lawson

Manufacture Date 31-Jul-2020

Follow all storage requirements, keep tightly closed when not in use and use good laboratory practices when handling.

*This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.*

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.