



2040 Savage Road
Charleston, South Carolina 29407
Phone (843) 763-4884
Fax (843) 766-9182
www.o2si.com

Quality System
Audited & Registered
by NSF-ISR to ISO 9001:2000

Date Received: _____


Certificate of Analysis

Page 1

Catalog No.: SR00500 Lot No.: 132155 Storage: Freeze Solvent: Isooctane Exp. Date: 2/4/2011 Description: Custom Cholronitrobenzenes Mix, 5-02, 100 mg/L, 2 x 5 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,3-trichloro-4-nitrobenzene	17700-09-3	99	2075.5.1P	99.04 +/- 1.38
1,2,3-trichloro-5-nitrobenzene	20098-48-0	98	2076.9.1P	99.57 +/- 1.39
1,3,5-trichloro-2-nitrobenzene	18708-70-8	98	2077.29.1P	99.86 +/- 1.39
2,3,4,5-tetrachloronitrobenzene	879-39-0	98	2074.7.1P	100.1 +/- 1.40
pentachloronitrobenzene	82-68-8	99	95.1.1P	100 +/- 1.4

Certified By: _____


Mike Wrigley

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.