



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

INVOICE #: CS314100
PO #: POR19840

CATALOG #: F2516JS

DESCRIPTION: 5-alpha-Androstane Solution

LOT #: 429-134B

EXPIRATION DATE: 04/10

Chem Service, Inc guarantees the concentration of F2516JS is 2000ug/ml \pm 5% in Methylene chloride

This product was compared to an independently prepared sample.
Determined by either GC/FID, GC/TCD, GC/ECD, GC/MS or HPLC.
This data is available upon request.

Weights of analytes less than or equal to 99% are corrected for impurities.

Our standards are suitable for use with all EPA methods.

Certified By:

John Conrad
CSM/TC



ISO 9001
Certificate Number: 31610



ISO/IEC 17025
Accreditation Number: 63520