

CERTIFICATE OF ANALYSIS

General Product Data:

Product Name: Dibenzo[a,i]pyrene-d₁₄

Product No.: 1529.24-10-DC/AN

CAS No.: [158776-07-9]

Batch Specific Data:

Batch No.: 7007

Product description:

Analyte: Dibenzo[a,i]pyrene-d₁₄, 98.7 % (NMR), Batch 6013
Solvent: Acetonitrile, Ultra resi-analyzed, 99.8%, lot 0707400027
Dichlormethane, Ultra resi-analyzed, 99.8%, lot E08465

Concentration: 0.01 mg/mL
The concentration has not been adjusted for compound impurities

Volume: 1.1 mL

Tolerance: The uncertainty in the preparation of this standard is less than $\pm 5\%$

Appearance: Clear colourless liquid

Storage: Dark and cool
Before opening allow the standard to reach room temperature

Expiry date: Guaranteed 3 years from date of issue

Trondheim, 31 January 2008

Issued by:

Ellen S. Broholm

Ellen S. Broholm

Approved by:

Inge Fenstad
Inge Fenstad



Chiron AS

Stiklestadvn. 1
N-7041 Trondheim
Norway

Phone No.:
+47 73 87 44 90

Fax No.:
+47 73 87 44 99

E-mail:
chiron@chiron.no

Website:
www.chiron.no

Org. No.:
967 607 657