



7290B Investment Drive North Charleston, South Carolina 29417
 Phone: 866.272.0932 Fax: 866.509.5146 www.o2si.com



ISO 17025 Accredited Chemical Testing Lab
 Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
020054-01	372301	≤ -10 °C	P/T Methanol	21-Mar-2021	1,2,4-Trichlorobenzene Solution, 1,000 mg/L, 5 x 1 mL
-5PAK					

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1003 ± 10.47 mg/L

Certified By: _____

Kara Catron

Manufacture Date 22-Mar-2019

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling. This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically. The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.