

Date Received: \_\_\_\_\_

**Certificate of Analysis** Rev 0 Page 1 of 1

**Catalog No.:** 020132-02    **Lot No.:** 335361    **Storage:** ≤ -10 °C    **Solvent:** P/T Methanol    **Exp. Date:** 24-Jul-2020    **Description:** Fluorobenzene Solution, 2,000 mg/L, 1 mL

<u>Compound</u>	<u>CAS No.</u>	<u>Purity (%)</u>	<u>Neat Material Lot No.</u>	<u>Concentration</u>
fluorobenzene	462-06-6	99.9	132.7.2P	2024 +/- 29.57 mg/L

Certified By: \_\_\_\_\_



Katrina Emelianoff

Manufacture Date 20-Jul-2017

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling. This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically. The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.