

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: 110182-09	Lot No.: 325271	Storage: ≤ -10 °C	Solvent: Methylene Chloride	Exp. Date: 1-Nov-2019	Description: GPC Calibration Solution, 5-1, Various Concentrations, 200, 800, 2000, 10000, & 250000 mg/L, 10 x 1 ml
-10PAK					

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	10010 ± 105.95 mg/L
corn oil	8001-30-7	100	507.1001.1P	250100 ± 866.37 mg/L
sulfur	7704-34-9	99	938.29.1P	795.9 ± 3.4 mg/L
methoxychlor	72-43-5	94.3	241.1.1P	1995 ± 21.19 mg/L
perylene	198-55-0	99.6	939.1.2.3P	199.2 ± 2.11 mg/L



Certified By: _____

Katrina Emelianoff
 Manufacture Date 27-Oct-2016

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.
 This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.
 The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.