

QUALITY CONTROL CERTIFICATE



The enclosed liquid handling instrument has been manufactured in a facility that employs a registered and certified quality management system. This system ensures that the development and production of this product has been strictly controlled and that it meets the highest quality standards.

Each instrument passes stringent validation and Control Procedures based on **EN ISO 8655**.

Model:	ME-200	Volume (μ l) :	20.000	Volume (μ l) :	200.00
Serial:	062130010				
Cat.No.:	14.1.2	Accuracy (%) :	0.722	Accuracy (%) :	0.474
Date:	17.11.10	Precision (%) :	0.488	Precision (%) :	0.099
Tested by:	C10				
Status:		Passed		Passed	

Calibrated and tested using the gravimetric method with distilled water:

Condition of measurements:

Basis of adjustment: Ex; Reference temperature: 20°C; Relative air humidity: 50%; Barometric pressure: 101kPa