



Quality System
Audited & Registered
by NSF-ISR to ISO 9001:2008

Certificate of Analysis

Rev 0

Catalog No: 160222-04-03-PUG
Lot No: 1094310
Expiration Date: 1-Aug-2017
Storage: Ambient
pH: 4.00@25°C
Description: Colorless pH Buffer Solution 4.00 ± 0.01 at 25°C,
500 mL

Date Received: _____

This standard was prepared gravimetrically using balances calibrated with NIST traceable weights (NIST Test Number 822/264157-00). Only calibrated Class A volumetric glassware was used to prepare this standard. Sub-boiled distilled acid and 18 megaohm deionized water were used to stabilize the product. All raw materials were checked for stoichiometry and purity prior to use. This standard has been spectrometrically certified by an independent source, which is directly traceable to NIST.

Chelsea Smith

Certified By: _____

Chelsea Smith

2030 Savage Road • Charleston, SC 29407
Phone: 866.272.0932 • Fax: 866.509.5146
www.o2si.com