

Certificate of Analysis

Rev 0



Quality System Audited & Registered
by NSF-ISR to ISO 9001:2008

Catalog No: AA-060020-07-125
Lot No: 1057439
Storage: Ambient
Matrix: 1% HNO₃
Container: 125 ml Narrow Mouth, HDPE clean



ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received _____
Expiration Date 1/Apr/2015
Certification Date 4/Mar/2014

Description: AA Grade Calcium, 125 mL
100 mg/L \pm 2% in 1% HNO₃

Element	Symbol	CAS No.	Source Lot No.	Purity %	Concentration mg/L
Calcium	Ca	7440-70-2	7006.29.2P	99.999	100 +/-2

The concentration of this standard was calculated based on the weight and volumes used in the manufacturing process. Sub-boiled distilled acid and 18 megaohm deionized water were used to stabilize the product. All raw materials were checked for stoichiometry and purity prior to use. This standard has been verified by Inductively Coupled Plasma Spectroscopy (ICP) and is traceable to NIST.

Glassware Calibration - Only calibrated Class A volumetric glassware was used to prepare this standard using Standard Operating Procedure (SOP) O2-LB-006.

Balance Calibration - This standard was prepared gravimetrically using balances calibrated annually by an A2LA accredited laboratory and analytical weights traceable to NIST (NIST Test Number 822/264157-00). All balances are checked daily using in-house procedure O2-LB-004.

Certified By: _____

Zachary Spinelli

2030 Savage Road • Charleston, SC 29407
Phone: 866.272.0932 • Fax: 866.509.5146
www.o2si.com