o2si smart solutions®

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
021354-03 -5PAK	262062	≤ -10 Degrees C	P/T Methanol	2-Oct-2017	3-Chlorotoluene Solution, 100 mg/L, 5 x 1 ml

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
3-chlorotoluene	108-41-8	100	1354.1.1P	100.4 +/- 1.45	mg/L

Certified By: Amara Q Iverson

Amara Iverson

Manufacture Date 3-Oct-2015

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.