



Quality System Audited & Registered  
by NSF-ISR to ISO 9001:2008

2030 Savage Road  
Charleston, South Carolina 29407  
Phone: 866.272.0932  
Fax: 866.509.5146  
www.o2si.com



ISO 17025 Accredited Chemical Testing Lab  
Cert. No. 303 1.01

Date Received: \_\_\_\_\_

# Certificate of Analysis

Rev 0

Page 1 of 1

<b>Catalog No.:</b> 020135-30	<b>Lot No.:</b> 371359	<b>Storage:</b> ≤ -10 °C	<b>Solvent:</b> P/T Methanol	<b>Exp. Date:</b> 18-Mar-2022	<b>Description:</b> 4-Bromofluorobenzene Solution, 5,000 mg/L, 5 x 1 ml
<b>-5PAK</b>					

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
4-bromofluorobenzene (BFB)	460-00-4	99.5	135.7.1P	5047 +/- 10.86 mg/L

Certified By: \_\_\_\_\_

Kara Catron

Manufacture Date 14-Mar-2019

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.  
*This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.*

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.  
The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.