



2030 Savage Road
 Charleston, South Carolina 29407
 Phone: 866.272.0932
 Fax: 866.509.5146
 www.o2si.com



ISO 17025 Accredited Chemical Testing Lab
 Cert. No. 3031.01

Date Received: _____

Certificate of Analysis Rev 0 Page 1 of 1

Catalog No.: 030669-06	Lot No.: 376019	Storage: ≤ -10 °C	Solvent: Methanol	Exp. Date: 10-May-2022	Description: Dicamba Solution, 50 mg/L, 2 x 1 ml
-2PAK					

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
Dicamba	1918-00-9	97.2	669.3.4P	49.5 +/- 1.15 mg/L

Certified By: _____

Joanna Radu
 Manufacture Date 6-May-2019

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.
This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.
 The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.